Notice of Allowability	Application No.	Applicant(s)
	10/618,571	CONNORS ET AL.
	Examiner	Art Unit
	John P. Lacyk	3736
The MAILING DATE of this communication appears on the cover sheet with the correspondence address All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS. This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.		
1. This communication is responsive to request for RCE filed 05/31/05.		
2. The allowed claim(s) is/are <u>1-46</u> .		
3. The drawings filed on 11 July 2003 are accepted by the Examiner.		
4.		
Attachment(s) 1. ☐ Notice of References Cited (PTO-892) 2. ☐ Notice of Draftperson's Patent Drawing Review (PTO-948) 3. ☑ Information Disclosure Statements (PTO-1449 or PTO/SB/0 Paper No./Mail Date 05/31/05 4. ☐ Examiner's Comment Regarding Requirement for Deposit of Biological Material	6. ☐ Interview Summary Paper No./Mail Da 08), 7. ☑ Examiner's Amendr	te

1. A request for continued examination under 37 CFR 1.114, including the fee set forth in 37 CFR 1.17(e), was filed in this application after allowance or after an Office action under *Ex Parte Quayle*, 25 USPQ 74, 453 O.G. 213 (Comm'r Pat. 1935). Since this application is eligible for continued examination under 37 CFR 1.114, and the fee set forth in 37 CFR 1.17(e) has been timely paid, prosecution in this application has been reopened pursuant to 37 CFR 1.114. Applicant's submission filed on 31 May 2005 has been entered.

Page 2

2. An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR 1.312. To ensure consideration of such an amendment, it MUST be submitted no later than the payment of the issue fee.

The application has been amended as follows:

In paragraph 0001 of the specification:

On line 3: after November 27, 2000, - - now U.S. Patent No. 6,682,473 - - has been added.

3. The following is an examiner's statement of reasons for allowance: The prior art fails to show an attenuation device and method having a flexible housing configured to be introduced into a bladder in a first configuration and then at least partially inflated into a second configuration; and having a high vapor pressure media with a vapor pressure

Page 3

Art Unit: 3736

approximately equal to the intravesical pressure of the bladder and a permeability of less than about 1 ml/day at body temperature through the outer wall of the flexible housing.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Any inquiry concerning this communication or earlier communications from the examiner should be directed to John P. Lacyk whose telephone number is 571-272-4728. The examiner can normally be reached on Mon-Fri, 8:30-5:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Max Hindenburg can be reached on 571-272-4726. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Application/Control Number: 10/618,571 Page 4

Art Unit: 3736

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

John P Lacyk
Primary Examiner
Art Unit 3736

J.P. Lacyk